

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025			Page 1 of 13
Registration No 572			
Nominal calibration temperature 22°C ± 3°C. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of k=2. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.			
Class No	Range	Least uncertainty	
5.28	Flow Measuring Devices		
	(a) Anemometers (Vane & Hotwire)		
	0.1 to 2.6m/sec	± 0.06m/sec or 1% of reading whichever is	
	2.6 to 15m/sec	± 0.06m/sec or 1% of reading whichever is	
	(l) Pitot tubes		
	Standard: 1.0 to 15m/sec	± 0.07m/sec or 1.6% of reading whichever is	
	Type S: 1.0 to 15m/sec	± 0.07m/sec or 1.6% of reading whichever is	
5.31	Volumetric Equipment		
	(b) Air samplers relating to (ISO/CD 14698-1, page 16 under B.3.3.2, 1996)		
	Nominal 100 Litres/min	± 1.1 Litres	
5.42	Differential Pressure Devices		
	(a), (b), (c) Diaphragm types, Liquid column types (inclined & vertical) & other		
	0.001 to 2 inches H ₂ O	± 0.001"	
	2 to 4 inches H ₂ O	± 0.001"	
	4 to 6 inches H ₂ O	± 0.001"	
	6 to 8 inches H ₂ O	± 0.002"	
	8 to 10 inches H ₂ O	± 0.002"	
	10 to 12 inches H ₂ O	± 0.002"	
	12 to 14 inches H ₂ O	± 0.003"	
	14 to 16 inches H ₂ O	± 0.003"	
	16 to 18 inches H ₂ O	± 0.003"	
	18 to 20 inches H ₂ O	± 0.004"	
	20 to 22 inches H ₂ O	± 0.004"	
	22 to 24 inches H ₂ O	± 0.004"	

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 2 of 13
 Registration No 572

Nominal calibration temperature $22^{\circ}\text{C} \pm 3^{\circ}\text{C}$. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.61	Temperature Measuring Equipment	
	(a) Indicators, calibrators, recorders & controllers by simulated rare metal thermocouple output using multifunction calibration standard.	
	Type B +600 to +800°C	$\pm 0.5^{\circ}\text{C}$
	+800 to +1000°C	$\pm 0.4^{\circ}\text{C}$
	+1000 to +1550°C	$\pm 0.2^{\circ}\text{C}$
	+1550 to +1820°C	$\pm 0.2^{\circ}\text{C}$
	Type R 0 to +250°C	$\pm 0.51^{\circ}\text{C}$
	+250 to +400°C	$\pm 0.51^{\circ}\text{C}$
	+400 to +1000°C	$\pm 0.51^{\circ}\text{C}$
	+1000 to +1767°C	$\pm 0.51^{\circ}\text{C}$
	Type S 0 to +250°C	$\pm 0.51^{\circ}\text{C}$
	+250 to +1000°C	$\pm 0.51^{\circ}\text{C}$
	+1000 to +1400°C	$\pm 0.51^{\circ}\text{C}$
	+1400 to +1767°C	$\pm 0.51^{\circ}\text{C}$
5.6.1	(b) Indicators, calibrators, recorders & controllers by simulated base metal thermocouple output using multifunction calibration standard.	
	Type E -250 to -100°C	$\pm 0.07^{\circ}\text{C}$
	-100 to -25°C	$\pm 0.06^{\circ}\text{C}$
	-25 to +350°C	$\pm 0.06^{\circ}\text{C}$
	+350 to +650°C	$\pm 0.06^{\circ}\text{C}$
	+650 to +1000°C	$\pm 0.06^{\circ}\text{C}$
	Type J -210 to -100°C	$\pm 0.06^{\circ}\text{C}$
	-100 to -30°C	$\pm 0.06^{\circ}\text{C}$
	-30 to +150°C	$\pm 0.06^{\circ}\text{C}$
	+150 to +760°C	$\pm 0.06^{\circ}\text{C}$
	+760 to +1200°C	$\pm 0.06^{\circ}\text{C}$

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 3 of 13
 Registration No 572

Nominal calibration temperature $22^{\circ}\text{C} \pm 3^{\circ}\text{C}$. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.61	Temperature Measuring Equipment	
	(b) Indicators, calibrators, recorders & controllers by simulated base metal thermocouple output using multifunction calibration standard.	
	Type K -200 to -100°C	$\pm 0.09^{\circ}\text{C}$
	-100 to -25°C	$\pm 0.07^{\circ}\text{C}$
	-25 to +120°C	$\pm 0.07^{\circ}\text{C}$
	+120 to +1000°C	$\pm 0.08^{\circ}\text{C}$
	+1000 to +1372°C	$\pm 0.08^{\circ}\text{C}$
	Type N -200 to -100°C	$\pm 0.12^{\circ}\text{C}$
	-100 to -25°C	$\pm 0.09^{\circ}\text{C}$
	-25 to +120°C	$\pm 0.08^{\circ}\text{C}$
	+120 to +410°C	$\pm 0.08^{\circ}\text{C}$
	+410 to +1300°C	$\pm 0.09^{\circ}\text{C}$
	Type T -250 to -150°C	$\pm 0.12^{\circ}\text{C}$
	-150 to 0°C	$\pm 0.08^{\circ}\text{C}$
	0 to +120°C	$\pm 0.07^{\circ}\text{C}$
	+120 to +400°C	$\pm 0.06^{\circ}\text{C}$

TR Corporation NZ LTD

Class No	Range	Least uncertainty
SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 4 of 13 Registration No 572		
Nominal calibration temperature 22°C ± 3°C. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of k=2. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.		
5.61	Temperature Measuring Equipment	
	(c) Indicators, calibrators, recorders & controllers by simulated platinum resistance thermometer output using multifunction calibration standard.	
Pt385 100Ω 3W or 4W ITS- 90 or IPTS-68	-200 to -80°C	± 0.003°C
	-80 to 0°C	± 0.008°C
	0 to +100°C	± 0.011°C
	+100 to +300	± 0.02°C
	+300 to +400°C	± 0.02°C
	+400 to +630°C	± 0.03°C
	+630 to +800°C	± 0.04°C
5.61	(a-e,k-p) Indicators, recorders, controllers & transmitters, including direct reading temperature measuring systems by direct comparison using stirred liquid bath in conjunction with a reference thermometer.	
	-25 to 0°C	± 0.095°C
	0 to +70°C	± 0.095°C
	+70 to +140°C	± 0.095°C
5.89	Indicating instruments and Recording Instruments	
	(a) DC Voltmeters	
	± 0 to 330mV dc	± 16ppm
	± 0 to 3.3V dc	± 13ppm
	± 0 to 33V dc	± 17ppm
	± 30 to 330V dc	± 17ppm
	± 100 to 1020V dc	± 17ppm

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 5 of 13
Registration No 572

Nominal calibration temperature 22°C ± 3°C. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.89	Indicating Instruments and Recording Instruments	
	(b) AC Voltmeters	
	1.0 to 33mV 10 to 45Hz	± 0.05%
	45Hz to 10kHz	± 0.05%
	10 to 20kHz	± 0.05%
	20 to 50kHz	± 0.06%
	50 to 100kHz	± 0.21%
	100 to 500kHz	± 0.36%
	33 to 330mV 10 to 45Hz	± 0.013%
	45Hz to 10kHz	± 0.013%
	10 to 20kHz	± 0.013%
	20 to 50kHz	± 0.024%
	50 to 100kHz	± 0.06%
	100 to 500kHz	± 0.29%
	0.33 to 3.3V 10 to 45Hz	± 0.023%
	45Hz to 10kHz	± 0.013%
	10 to 20kHz	± 0.013%
	20 to 50kHz	± 0.024%
	50 to 100kHz	± 0.048%
	100 to 500kHz	± 0.35%
	3.3 to 33V 10 to 45Hz	± 0.021%
	45Hz to 10kHz	± 0.015%
	10 to 20kHz	± 0.015%
	20 to 50kHz	± 0.043%
	50 to 100kHz	± 0.1%
	33 to 330V 45Hz to 1kHz	± 0.015%
	1 to 10kHz	± 0.025%
	10 to 20kHz	± 0.046%
	330 to 1000V 45Hz to 1kHz	± 0.02%
	1 to 5kHz	± 0.02%
	5 to 10kHz	± 0.02%

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 6 of 13
 Registration No 572

Nominal calibration temperature 22°C ± 3°C. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.89	Indicating Instruments and Recording Instruments	
	(c) DC Ammeters	
	± 0 to 3.3mA	± 118ppm
	± 0 to 33mA	± 116ppm
	± 0 to 330mA	± 234ppm
	± 0 to 2.2A	± 110ppm
	± 0 to 11A	± 184ppm
5.89	(c) DC Ammeters Clamp on type	
	± 11 to 16.5A	± 0.57A
	± 16.5 to 110A	± 0.57A
	± 110 to 550A	± 0.57A
5.89	(d) AC Ammeters	
Sine Wave	29 to 330µA 10 to 20Hz	± 0.53%
	20 to 45Hz	± 0.29%
	45 to 1kHz	± 0.16%
	1 to 5kHz	± 0.3%
	5 to 10kHz	± 0.3%
	0.33 to 3.3mA 10 to 20Hz	± 0.53%
	20 to 45Hz	± 0.24%
	45 to 1kHz	± 0.11%
	1 to 5kHz	± 0.25%
	5 to 10kHz	± 0.51%
	3.3 to 33mA 10 to 20Hz	± 0.53%
	20 to 45Hz	± 0.24%
	45 to 1kHz	± 0.11%
	1 to 5kHz	± 0.25%
	5 to 10kHz	± 0.3%
	33 to 330mA 10 to 20Hz	± 0.51%
	20 to 45Hz	± 0.24%
	45 to 1kHz	± 0.11%
	1 to 5kHz	± 0.25%
	5 to 10kHz	± 0.51%

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 7 of 13
Registration No 572

Nominal calibration temperature $22^{\circ}\text{C} \pm 3^{\circ}\text{C}$. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.89	Indicating Instruments and Recording Instruments	
Sine Wave	(d) AC Ammeters	
	0.33 to 2.2A 10 to 45Hz	$\pm 0.15\%$
	45Hz to 1kHz	$\pm 0.08\%$
	1 to 5kHz	$\pm 0.09\%$
	2.2 to 11A 45 to 65Hz	$\pm 0.12\%$
	65 to 500Hz	$\pm 0.12\%$
	500Hz to 1kHz	$\pm 0.12\%$
5.89	(d) AC Ammeters Clamp on Type	
	11 to 16.5A 10 to 20Hz	$\pm 0.57\text{A}$
	20 to 45Hz	$\pm 0.57\text{A}$
	45Hz to 1kHz	$\pm 0.57\text{A}$
	1 to 5kHz	$\pm 0.57\text{A}$
	5 to 10kHz	$\pm 0.57\text{A}$
	16.5 to 110A 10 to 45Hz	$\pm 0.57\text{A}$
	45Hz to 1kHz	$\pm 0.57\text{A}$
	1 to 5kHz	$\pm 0.57\text{A}$
	110 to 550A 45 to 65Hz	$\pm 0.57\text{A}$
	65 to 500Hz	$\pm 0.57\text{A}$
	500Hz to 1kHz	$\pm 0.57\text{A}$
	5.89	(i) Ohmmeters
2W & 4W	0 to 11Ω	$\pm 52\text{ppm}$
	11 to 33Ω	$\pm 46\text{ppm}$
	33 to 110Ω	$\pm 35\text{ppm}$
	110 to 330Ω	$\pm 28\text{ppm}$
	330 to $1.1\text{k}\Omega$	$\pm 23\text{ppm}$
	1.1 to $3.3\text{k}\Omega$	$\pm 23\text{ppm}$
	3.3 to $11\text{k}\Omega$	$\pm 28\text{ppm}$
	11 to $33\text{k}\Omega$	$\pm 33\text{ppm}$
	33 to $110\text{k}\Omega$	$\pm 35\text{ppm}$
110 to $330\text{k}\Omega$	$\pm 38\text{ppm}$	

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 8 of 13
Registration No 572

Nominal calibration temperature $22^{\circ}\text{C} \pm 3^{\circ}\text{C}$. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.89	(i) Ohmmeters	
	330k to 1.1M Ω	$\pm 47\text{ppm}$
	1.1 to 3.3M Ω	$\pm 90\text{ppm}$
	3.3 to 11M Ω	$\pm 145\text{ppm}$
	11 to 33M Ω	$\pm 0.43\%$
	33 to 110M Ω	$\pm 0.06\%$
	110 to 330M Ω	$\pm 0.29\%$
5.88	Calibrators for Instrumentation (source)	
	(a) DC voltage	
	± 0 to 100mV	$\pm 23\text{ppm}$
	± 0 to 1V	$\pm 12\text{ppm}$
	± 0 to 10V	$\pm 11\text{ppm}$
	± 0 to 100V	$\pm 12\text{ppm}$
	± 0 to 1000V	$\pm 14\text{ppm}$
True RMS Sine Wave	(b) AC voltage	
	0 to 100mV 3Hz to 10Hz	$\pm 247\text{ppm}$
	10Hz to 20kHz	$\pm 247\text{ppm}$
	20 to 50kHz	$\pm 0.05\%$
	0 to 1V 3Hz to 10Hz	$\pm 110\text{ppm}$
	10Hz to 20kHz	$\pm 110\text{ppm}$
	20 to 50kHz	$\pm 186\text{ppm}$
	0 to 10V 3Hz to 10Hz	$\pm 112\text{ppm}$
	10Hz to 20kHz	$\pm 112\text{ppm}$
	20 to 50kHz	$\pm 198\text{ppm}$
	0 to 100V 3Hz to 10Hz	$\pm 121\text{ppm}$
	10Hz to 20kHz	$\pm 121\text{ppm}$
	20 to 50kHz	$\pm 334\text{ppm}$
	0 to 750V 3Hz to 10Hz	$\pm 117\text{ppm}$
	10Hz to 20kHz	$\pm 117\text{ppm}$
	20 to 50kHz	$\pm 171\text{ppm}$

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 9 of 13
 Registration No 572

Nominal calibration temperature $22^{\circ}\text{C} \pm 3^{\circ}\text{C}$. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.88	Calibrators for Instrumentation (source)	
	(c) DC current	
	± 0 to 10mA	$\pm 84\text{ppm}$
	± 0 to 100mA	$\pm 90\text{ppm}$
	± 0 to 1A	$\pm 145\text{ppm}$
	± 0 to 3A	$\pm 155\text{ppm}$
5.88	(d) AC current	
True RMS Sine Wave	0 to 1A <i>3Hz to 10Hz</i>	$\pm 0.1\%$
	<i>10Hz to 5kHz</i>	$\pm 0.1\%$
	0 to 3A <i>3Hz to 10Hz</i>	$\pm 0.11\%$
	<i>10Hz to 5kHz</i>	$\pm 0.11\%$
5.88	(e) Resistance	
	Ohms 4W 0.0001 Ω to 100 Ω	$\pm 0.002\Omega$
	<i>0.001Ω to 1kΩ</i>	$\pm 0.017\Omega$
	<i>0.01Ω to 10kΩ</i>	$\pm 0.17\Omega$
	<i>0.1Ω to 100kΩ</i>	$\pm 2.7\Omega$
	Ohms 2W 1.0 Ω to 1M Ω	$\pm 38\Omega$
	<i>10Ω to 10MΩ</i>	$\pm 640\Omega$
	<i>100Ω to 100MΩ</i>	$\pm 40.2\text{k}\Omega$

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 10 of 13
Registration No 572

Nominal calibration temperature $22^{\circ}\text{C} \pm 3^{\circ}\text{C}$. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.91	Frequency Measurement and Time Measurement	
	(a) Frequency Meters	
	Sine Wave. AC Input range 1mV to 3.3V	
	0.01 to 9.99Hz	$\pm 0.005\text{mHz}$
	Sine Wave AC Input range 0.33 to 3.29999V	
	10 to 119.99Hz	$\pm 0.05\text{mHz}$
	120 to 1199Hz	$\pm 0.5\text{mHz}$
	1.2 to 9.999kHz	$\pm 0.01\text{mHz}$
	10 to 11.999kHz	$\pm 0.05\text{mHz}$
	12 to 119.99kHz	$\pm 0.5\text{mHz}$
	120 to 1199.9kHz	$\pm 0.05\text{mHz}$
	1.2 to 2.0MHz	$\pm 0.05\text{mHz}$
	10MHz (fixed sine wave)	$\pm 0.99999\text{mHz}$
5.91	(c) Counters (<i>Period Measurement</i>)	
	Square Wave. AC Input range 1mV to 3.3V	50% duty cycle
	0.1s to 100s	$\pm 0.8\%$ of period + 100ns
	83ms to 0.1s	$\pm 0.8\%$ of period + 100ns
	0.83ms to 83ms	$\pm 0.8\%$ of period + 100ns
	83 μs to 0.83ms	$\pm 0.8\%$ of period + 100ns
	0.83 μs to 83 μs	$\pm 0.8\%$ of period + 100ns
	0.83 μs to 8.3 μs	$\pm 0.8\%$ of period + 100ns
	0.5 μs to 0.83 μs	$\pm 0.8\%$ of period + 100ns
5.93	Signal Sources	
	(e) Other (Pt Pt385 100Ω 3W or 4W IEC751-1995)	
	-200 to -80°C	$\pm 0.0048^{\circ}\text{C}$
	-80 to 0°C	$\pm 0.0052^{\circ}\text{C}$
@1mA	0 to $+100^{\circ}\text{C}$	$\pm 0.044^{\circ}\text{C}$
	+100 to +300	$\pm 0.046^{\circ}\text{C}$
	+300 to $+400^{\circ}\text{C}$	$\pm 0.048^{\circ}\text{C}$
	+400 to $+630^{\circ}\text{C}$	$\pm 0.051^{\circ}\text{C}$

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 11 of 13
Registration No 572

Nominal calibration temperature $22^{\circ}\text{C} \pm 3^{\circ}\text{C}$. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.93	Signal Sources	
	(e) Other (Rare Metal Thermocouple)	
Type B	+600 to +800°C	$\pm 0.5^{\circ}\text{C}$
	+800 to +1000°C	$\pm 0.4^{\circ}\text{C}$
	+1000 to +1550°C	$\pm 0.2^{\circ}\text{C}$
Type R	0 to +250°C	$\pm 0.51^{\circ}\text{C}$
	+250 to +400°C	$\pm 0.51^{\circ}\text{C}$
	+400 to +1000°C	$\pm 0.51^{\circ}\text{C}$
	+1000 to +1767°C	$\pm 0.51^{\circ}\text{C}$
Type S	0 to +250°C	$\pm 0.51^{\circ}\text{C}$
	+250 to +1000°C	$\pm 0.51^{\circ}\text{C}$
	+1000 to +1400°C	$\pm 0.51^{\circ}\text{C}$
	+1400 to +1767°C	$\pm 0.51^{\circ}\text{C}$
5.6.1	(e) Other (Base Metal Thermocouple)	
Type E	-250 to -100°C	$\pm 0.07^{\circ}\text{C}$
	-100 to -25°C	$\pm 0.06^{\circ}\text{C}$
	-25 to +350°C	$\pm 0.06^{\circ}\text{C}$
	+350 to +650°C	$\pm 0.06^{\circ}\text{C}$
	+650 to +1000°C	$\pm 0.06^{\circ}\text{C}$
Type J	-210 to -100°C	$\pm 0.06^{\circ}\text{C}$
	-100 to -30°C	$\pm 0.06^{\circ}\text{C}$
	-30 to +150°C	$\pm 0.06^{\circ}\text{C}$
	+150 to +760°C	$\pm 0.06^{\circ}\text{C}$
	+760 to +1200°C	$\pm 0.06^{\circ}\text{C}$

TR Corporation NZ LTD

SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 12 of 13
Registration No 572

Nominal calibration temperature $22^{\circ}\text{C} \pm 3^{\circ}\text{C}$. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.

Class No	Range	Least uncertainty
5.93	Signal Sources	
	(e) Other (Base Metal Thermocouple)	
Type K	-200 to -100°C	$\pm 0.09^{\circ}\text{C}$
	-100 to -25°C	$\pm 0.07^{\circ}\text{C}$
	-25 to +120°C	$\pm 0.07^{\circ}\text{C}$
	+120 to +1000°C	$\pm 0.08^{\circ}\text{C}$
	+1000 to +1372°C	$\pm 0.08^{\circ}\text{C}$
Type N	-200 to -100°C	$\pm 0.12^{\circ}\text{C}$
	-100 to -25°C	$\pm 0.09^{\circ}\text{C}$
	-25 to +120°C	$\pm 0.08^{\circ}\text{C}$
	+120 to +410°C	$\pm 0.08^{\circ}\text{C}$
	+410 to +1300°C	$\pm 0.09^{\circ}\text{C}$
Type T	-250 to -150°C	$\pm 0.12^{\circ}\text{C}$
	-150 to 0°C	$\pm 0.08^{\circ}\text{C}$
	0 to +120°C	$\pm 0.07^{\circ}\text{C}$
	+120 to +400°C	$\pm 0.06^{\circ}\text{C}$
5.93	Signal Sources	
	(e) Other (Signal Generators 100mV to 750mVac rms) <i>Non Endorsed</i>	
	For sources > 10% of ACV range except 100mV range. On 100mV range, frequency must be >10Hz if source voltage is <20mV.	
Square Wave	100mV 3Hz to 500kHz	$\pm 100\text{ppm}$
	1.0V 3Hz to 500kHz	$\pm 100\text{ppm}$
	10.0V 3Hz to 500kHz	$\pm 100\text{ppm}$
	100V 3Hz to 500kHz	$\pm 100\text{ppm}$
	750V 3Hz to 500kHz	$\pm 100\text{ppm}$

TR Corporation NZ LTD

Class No	Range	Least uncertainty
<p>SCHEDULE TO CERTIFICATE OF ACCREDITATION ISO 17025 Page 13 of 13 Registration No 572</p>		
<p>Nominal calibration temperature $22^{\circ}\text{C} \pm 3^{\circ}\text{C}$. All measurement uncertainties are based on a confidence interval of 95% and a coverage factor of $k=2$. Unless stated elsewhere in this schedule, calibrations are performed at the premises of the accredited laboratory.</p>		
5.93	Signal Sources	
Square or Sine Wave	(e) Other (Signal Generators 25mV to 250mVac rms) <i>Non Endorsed</i>	
	0.1Hz to 9.9999999MHz	$\pm 0.113\text{Hz}$
	1Hz to 99.999999MHz	$\pm 1.132\text{Hz}$
	10Hz to 99.99999MHz	$\pm 11.316\text{Hz}$
	100Hz to 99.9999MHz	$\pm 113.16\text{Hz}$
	1kHz to 99.999MHz	$\pm 1.132\text{kHz}$
	10kHz to 99.99MHz	$\pm 11.32\text{kHz}$
Square or Sine Wave	(e) Other (Signal Generators timers and period measurement 25mV to 250mVpp)	
	0.1S to 10^7S	$\pm 1.13^{-15}\text{S}$
	0.01S to 10^6S	$\pm 1.132^{-14}\text{S}$
	1mS to 10^5S	$\pm 1.1316^{-13}\text{S}$
	0.1mS to 10^4S	$\pm 1.1316^{-12}\text{S}$
	100nS to 10^3S	$\pm 1.132^{-11}\text{S}$
10nS to 10^2S	$\pm 1.132^{-10}\text{S}$	